

# EU CERTIFICATE OF CONFORMITY

**THERMIS, spol. s r.o.**  
**Mateří 14**  
**614 00 Brno**  
**IČO: 26883643**

**Product:** Standard pressure gauge, diameter 40,50,63 mm

**Type:** 301, 308, 304, 358

**Range:** -100-0 kPa ... 0-60 kPa ... 0-100 MPa

**Accuracy:** 1,6%, 2,5%

**Product is designed for:** pressure and vacuum measurement of liquids, vapour and gases

The subject matter described above is in accordance with the relevant European Union harmonization regulations and Government regulations of the Czech Republic:

- Directive of the European Parliament and of the Council **2014/35/EU**;
- Directive of the European Parliament and of the Council **2014/30/EU**;
- Government regulations nr. **117/2016Sb.**
- Government regulations nr. **118/2016 Sb.**

Harmonized standards / technical specifications on which bases we declare compliance:

- EN 837-1:1998
- EN 837-2:1998
- EN 837-3:1998

This declaration is made under the sole responsibility of the manufacturer  
Thermis, spol. s.r.o., Mateří 14, Brno.

Signed for and by name:

Place and date of issue:  
Brno, July 2018



měřicí a regulační technika

Mateří 14, 614 00 Brno

tel./fax: 545 226 177 - 80

IČ: 26883643 DIČ: CZ26883643

Ing. Josef Škorpík  
Managing Director

The conformity assessment is documented by a set technical documentantation stored at the manufacturer.