

EU CERTIFICATE OF CONFORMITY

THERMIS, spol. s r.o.
Mateří 14
614 00 Brno
IČO: 26883643

Product: Glycerine pressure gauge with bourdon pen diameter 63, 100, 160 mm

Type: 304G, 358G, 384G, 322G, 313G, 323G

Range: -100-0 kPa ... 0-60 kPa ... 0-160 MPa

Accuracy: 1,6%,

Product is designed for: pressure and vacuum measurement of liquids, vapour and gases

The subject matter described above is in accordance with the relevant European Union harmonization regulations and Government regulations of the Czech Republic:

- Directive of the European Parliament and of the Council **2014/35/EU**;
- Directive of the European Parliament and of the Council **2014/30/EU**;
- Government regulations nr. **117/2016 Sb.**
- Government regulations nr. **118/2016 Sb.**

Harmonized standards / technical specifications on which bases we declare compliance:

- EN 837-1:1998
- EN 837-2:1998
- EN 837-3:1998

This declaration is made under the sole responsibility of the manufacturer
Thermis, spol. s r.o., Mateří 14, Brno.

Signed for and by name:

Place and date of issue:
Brno, May 2020



měřicí a regulační technika
Mateří 14, 614 00 Brno ③
tel./fax: 545 226 177 - 80
IČ: 26883643 DIČ: CZ26883643
.....
Ing. Josef Škorpík
Managing Director

The conformity assessment is documented by a set technical documentantation stored at the manufacturer.