

EU CERTIFICATE OF CONFORMITY

THERMIS, spol. s r.o.
Mateří 14
614 00 Brno
IČO: 26883643

Product: Thermomanometer
Type: 3081/3082, diameter 80mm
Thermal range: 0-120°C /(0-150°C)
Pressure range: 0-4 bar, 6, 10, 16 bar
Connecting thread: G1/2
Accuracy: gauge 1,6%, thermometer 2,5%
Product is designed for: pressure and vacuum measurement of liquids, vapour and gases

The subject matter described above is in accordance with the relevant European Union harmonization regulations and Government regulations of the Czech Republic:

- Directive of the European Parliament and of the Council **2014/35/EU**;
- Directive of the European Parliament and of the Council **2014/30/EU**;
- Government regulations nr. **117/2003** Sb.
- Government regulations nr. **118/2006** Sb.

Harmonized standards / technical specifications on which bases we declare compliance:

- ČSN EN 837-1:1998
- ČSN EN 837-2:1998
- ČSN EN 837-3:1998
-

A test report was used for conformity assessment Nr. 23/03.

The conformity assessment is documented by a set technical documentantation stored at the manufacturer TP MŠ-036/03, ed.Nr.1, operating instructions, drawing documentation and control procedures Nr. 033-01133-..., 033-01134-... .

Signed for and by name:

Place and date of issue:
Brno, June 2022



měřicí a regulační technika
Mateří 14, 614 00 Brno ③
tel./fax: 545 226 177 - 80
IČ: 26883643 DIČ: CZ26883643
.....
Ing. Josef Škorpík
Managing Director